

Oxygen & Argon Works Ltd.	Spec.No.G-01.002	
Product Manufacturing Control	Issue: C	Date: 14.02.2008
GAS SPECIFICATION	Sheet 1 of 1	

OXYGEN - Medical Grade (Breathing) - 99.5%
(Meeting the EP requirements)

Typical purity/Impurity Limits:

Oxygen:	≥ 99.5%
Carbon Dioxide:	≤ 300 ppm
Carbon Monoxide:	≤ 5 ppm
Moisture:	≤ 67 ppm

This grade of oxygen is suitable for medical and pharmaceutical application and, as such, meets the requirement of the EP.

Above specification applies to both Oxygen Gas and Oxygen Liquid.

Quality Assurance - According to company QC procedure No. 11.01.

Manufacture - According to Company Works' Instructions No. 01.01, 01.02, 10.01.

Packing & Distribution - In high pressure gas cylinders, liquid DEWARS and Cryogenic road tankers according to customers' requirements.

Valve Outlet Thread - According to IS 637 part 3.

[3/4" BSP, 14 teeth/", Male, Cylindrical, right hand or specific "pin index"]

Hazardous Materials Codes			
	U.N. No.	Emergency Actions	Hazard Class
GAS	1072	2S	2.2, 5.1
LIQUID	1073	2P	2.2, 5.1



Originator : Meir Ben Ishay	Date: 14.02.2008	Signature:
-----------------------------	------------------	------------